

CE mark date: June 25, 2019

## **Declaration of Conformity (No. E2020006)**

We, Sun Medical Co., Ltd., ensure and declare that the products listed in the schedule A below meet the provisions of the Directive 93/42/EEC concerning medical devices. Our products listed in the schedule A below are also in conformity with the safety requirements laid down in Annex I of the Directive 93/42/EEC as amended and the declarations of compliance of our products listed in the schedule A are issued according to Annex II excl. section 4 of the Directive 93/42/EEC as amended on Medical Devices.

EC Certificate No.: G1 026583 0028

Our authorized representative in EU is: EMERGO EUROPE B.V. having its principal office at Prinsessegracht 20, 2514 AP, The Hague, The Netherlands.

The notified body is: TÜV SÜD Product Service GmbH (ID No. 0123) at Ridlerstrasse 65. 80339 Munich, Germany

Schedule A (Part 1)

| Product            | Type of<br>Product      | Classification<br>(Applied Rule of Annex) | Lot Number<br>From: | Standards Applied   |   | Additional Information |
|--------------------|-------------------------|---|---------------------|---|---|------------------------|
| Hybrid Bond<br>ONE | Dental Bonding<br>Agent | IIa<br>(Annex IX, Rule 8)                 | In Part 2           | EN ISO 15223-1:2016<br>EN 1041:2008<br>EN 1641:2009<br>EN ISO 7405:2008 | EN ISO 10993-1:2009<br>EN ISO 10993-18:2009<br>EN ISO 14971:2012<br>EN 62366:2008 | -                      |

We are exclusively responsible for this statement.

Narimichi Honda

Management Representative for Quality Management System

M. Hond July, 2, 2020

Sun Medical Co., Ltd.

571-2 Furutaka-cho, Moriyama, Shiga, 524-0044 JAPAN

TEL: +81-77-582-9981

FAX: +81-77-582-9984



## Schedule A (Part 2)

| Product code | UDI-DI        | Product name    | Lot number<br>From: |
|--------------|---------------|-----------------|---------------------|
| T435E        | 4560227799782 | Hybrid Bond ONE | TV1                 |
| T435TR       | 4560227800655 | Hybrid Bond ONE |                     |